

United States Patent and Trademark Office

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.	
09/322,201	05/28/1999	SHIGERU TANAKA	236/238	7673	
7590 03/03/2004 U.S. SURGICAL CORPORATION Division of Tyco Healtcare Group LP 150 GLOVER AVENUE NORWALK, CT 06856			EXAMI	EXAMINER	
			DAWSON,	DAWSON, GLENN K	
			ART UNIT	PAPER NUMBER	
			3761	/	
,			DATE MAILED: 03/03/2004	1.8	

Please find below and/or attached an Office communication concerning this application or proceeding.

i					
	•	Application No.	Applicant(s)		
		09/322,201	TANAKA ET AL.		
	Office Action Summary	Examin r	Art Unit		
÷		Glenn K Dawson	3761		
Period fo	The MAILING DATE of this communication ap or Reply	opears on the cover sheet with the c	correspond nce address		
A SH THE - Exte after - If the - If NC - Failu Any	ORTENED STATUTORY PERIOD FOR REP MAILING DATE OF THIS COMMUNICATION nsions of time may be available under the provisions of 37 CFR 1 SIX (6) MONTHS from the mailing date of this communication. Period for reply specified above is less than thirty (30) days, a reduction of the provision of the pr	136(a). In no event, however, may a reply be tin ply within the statutory minimum of thirty (30) day d will apply and will expire SIX (6) MONTHS from tte, cause the application to become ABANDONE	nely filed s will be considered timely. the mailing date of this communication. D (35 U.S.C. § 133).		
Status					
1)⊠	Responsive to communication(s) filed on 16	December 2002 and 24 June 2003	3.		
	This action is FINAL. 2b)⊠ This action is non-final.				
3)	,—				
	closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213.				
Disposit	ion of Claims				
5)⊠ 6)⊠ 7)⊠	Claim(s) 1-6,8-39 ad 43-45 is/are pending in the application. 4a) Of the above claim(s) is/are withdrawn from consideration. Claim(s) 23-36 is/are allowed. Claim(s) 1-4,8-11,17-22,37,38 and 43-45 is/are rejected. Claim(s) 5,6,12-16 and 39 is/are objected to. Claim(s) are subject to restriction and/or election requirement.				
Applicat	ion Papers				
	The specification is objected to by the Examir	ner			
•	The drawing(s) filed on is/are: a) ac		Examiner.		
,,	Applicant may not request that any objection to th				
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).					
11)	The oath or declaration is objected to by the B	Examiner. Note the attached Office	Action or form PTO-152.		
Priority (under 35 U.S.C. § 119				
a)	Acknowledgment is made of a claim for foreign All b) Some * c) None of: 1. Certified copies of the priority documents. 2. Certified copies of the priority documents. 3. Copies of the certified copies of the priority application from the International Bure. See the attached detailed Office action for a list	nts have been received. nts have been received in Applicati fority documents have been receive au (PCT Rule 17.2(a)).	ion No ed in this National Stage		
Attachmen	t(s)				
1) Notice of References Cited (PTO-892) 4) Interview Summary (PTO-413)					
2) Notic 3) Infor	ce of Draftsperson's Patent Drawing Review (PTO-948) mation Disclosure Statement(s) (PTO-1449 or PTO/SB/06 er No(s)/Mail Date	Paper No(s)/Mail Da			

Art Unit: 3761

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-3,8-11,21,37 and 38 are rejected under 35 U.S.C. 102(b) as being anticipated by Hopper, et al.-5,697,946.

Hopper discloses a trocar cannula 22 having a tapered balloon 44 attached thereto. The balloon tapers from the proximal end to the distal end.

Claims 43-45 are rejected under 35 U.S.C. 102(b) as being anticipated by Fogarty, et al.-5,690,668.

Fogarty discloses a cannula 81 having a tapered balloon 92 attached thereto. The distal end as shown in fig. 5 would be of less diameter and volume than the proximal portion. The distal section of the balloon is inverted to shorten its length as shown in fig. 5 or 7.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Application/Control Number: 09/322,201

Art Unit: 3761

The factual inquiries set forth in *Graham* v. *John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

- 1. Determining the scope and contents of the prior art.
- 2. Ascertaining the differences between the prior art and the claims at issue.
- 3. Resolving the level of ordinary skill in the pertinent art.
- 4. Considering objective evidence present in the application indicating obviousness or nonobviousness.

This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

Claims 4,17-20 and 22 are rejected under 35 U.S.C. 103(a) as being unpatentable over Hopper, et al.-'946 in view of Kayan, et al.-5,324,270.

Hopper discloses the invention as claimed with the exception of the seal and the 2-part handle. Kayan discloses a seal, duckbill valve 37, and a 2 part handle 134,133, and even a 3rd part 81. It would have been obvious to have provided the cannula of Hopper with an internal sealing valve, as taught by Kayan, as this would allow for the insufflation of the body cavity. Using a 2-part handle on the instruments which are placed inside the cannula tube of Hopper would have been obvious as this allows for

the insertion and removal of various components through the cannula while not losing the insufflation gas from the body cavity.

Allowable Subject Matter

Claims 23-36 are allowed.

Claims 5,6,12-16 and 39 are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

Response to Arguments

Applicant's arguments with respect to the rejected claims have been considered but are most in view of the new ground(s) of rejection.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Glenn K Dawson whose telephone number is 703-308-4304. The examiner can normally be reached on M-Th 7:30-5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Michael J. Milano can be reached on 703-308-2496. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.



Art Unit: 3761

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Glenn K Dawson Primary Examiner Art Unit 3761

Gkd 27 February 2004